XENCO Laboratories
Trans West Analytical Services
Formerly Columbia Analytical Services

3725 E. Atlanta Avenue, Phoenix, AZ 85040 | 602-437-0330 | www.xenco.com

July 09, 2010

Chuck Gordon
Terranext
4050 E. Cotton Center Blvd.
Suite 73
Phoenix, AZ 85040

RE: WVBA/0310-3154
Work Order No.: 10060432

Dear Chuck,
XENCO Laboratories, Inc. received 2 samples on 6/28/10. The results of the analyses are presented in the following report.

The Case Narrative of this report addresses any Quality Control and/or Quality Assurance issues associated with this Work Order.

Analyses were performed according to our laboratory's NELAP-approved quality assurance program. The test results meet requirements of the current NELAP standards, where applicable, and except as noted in the laboratory case narrative provided. For a specific list of NELAP-accredited analytes, refer to the certifications section at www.xenco.com. All results are intended to be considered in their entirety and XENCO Laboratories is not responsible for use of less than the complete report. Results apply only to the items submitted to the laboratory for analysis and individual items (samples) analyzed, as listed in the report.

If you have any questions regarding these test results, please feel free to call us at:
(602) 437-0330.

Sincerely,

Skip Harden
Project Manager

ADHS License No. AZ0757/AZ0758/AZM758

XENCO Laboratories
Trans West Analytical Services
Formerly Columbia Analytical Services

Client: Terranext
Work Order: 10060432
Project Name: WVBA
Project Number: 0310-3154

Case Narrative

Samples were received intact and within proper temperature criteria.

All method blanks, laboratory spikes, and/or matrix spikes met quality control objectives for the parameters associated with this Work Order except as detailed below or on the Data Qualifier page of this report. Data Qualifiers used in this report are in accordance with ADEQ Arizona Data Qualifiers, Revision 3.0 9/20/2007.

Data qualifiers ("flags") contained within this analytical report have been issued to explain a quality control deficiency, and do not affect the quality (validity) of the data unless noted otherwise in the case narrative.

Trans West Analytical Services
Formerly Columbia Analytical Services

License No. AZ0757/AZM757

CLIENT: Terranext
Project Name: WVBA
Project Number: 0310-3154
Work Order: 10060432

Work Order Sample Summary

Client Sample ID	Lab Sample ID	Test Code	Collection Date	Date Received
AVB20-0302-1000	10060432-01A	SW8260B	6/28/10 09:49 AM	6/28/10 12:10 PM
AVB20-0300-14145	10060432-02A	SW8260B	6/28/10 09:49 AM	6/28/10 12:10 PM

Trans West Analytical Services
Formerly Columbia Analytical Services

License No. AZ0757/AZM757

CLIENT:	Terranext
Project Name:	WVBA
Project Number:	0310-3154
Work Order:	10060432
Date Received:	28-Jun-10

References

XENCO Laboratories, Inc. uses the methods outlined in the following references:

Code of Federal Regulations, 40CFR, Part 136, Appendix A, July 2005.

Standard Methods for the Examination of Water and Wastewater, 20th Edition, 1998.

Methods for Chemical Analysis of Water and Wastes, EPA-600/4-79-020, Revised March 1983.

Methods for the Determination of Inorganic Substances in Environmental Samples, EPA/600/R-93/100, Revised August 1993.

Methods for the Determination of Metals in Environmental Samples, Supplement 1: EPA/600/R-94/111, Revised May 1994.

Methods for the Determination of Organic Compounds in Drinking Water, EPA/600/4-88/039, Revised July, 1991; EPA-600/4-90/020, Supplement I, July 1990; EPA-600/R-92/129; Supplement II, August 1992; EPA-600/R-95/131, Supplement III, August 1995.

Hach, Water Analysis Handbook, 3rd Edition, 1997.

Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition, 1986 including Update I, July 1992; Update IIA, August 1993; Update II; September 1994; Update IIB, January 1995; Update III, December 1996. Update IIIA, June 1999; and Update IIIB July 2005.

Bureau of Laboratory Services, State of Arizona Department of Health Services Method 8015AZ.R1, September 1998. (Comment: C6-C10 GRO reported by this method is not to be used in compliance situations)

ASTM MethodD4982, Annual Book of ASTM Standards, Volumes 11.01 and 11.02, 1995

The Determination of Polychlorinated Biphenyls in Transformer Fluid and Waste Oils, EPA-600 4-81-045, September 1982.

EPA Method 9013A, Cyanide Extraction Procedure for Solids and Oils. (Rev, 1 November 2004)

EPA Method 5035A, Closed-System Purge-and-Trap and Extraction for Volatile Organics in Soil and Waste Samples (draft rev. 1 July 2002)

EPA Method 5030C, Purge-and-Trap for Aqueous Samples (rev.3 May 2003)

Office of Ground Water and Drinking Water Technical Support Center, EPA 815-R-05-004, Manual for Certification of Drinking Water, (5th Edition January 2005)

XENCO Laboratories

Trans West Analytical Services

Formerly Columbia Analytical Services

License No. AZ0757/AZM757

CLIENT: Terranext
Work Order: 10060432
Lab ID: 10060432-01
Project Name: WVBA
Project Number: 0310-3154

Client Sample ID: AVB20-0302-1000
Collection Date: 6/28/2010 9:49:00 AM
Matrix: Water

Analyte	Result	PQL	Qual	Units	DF	Date Prepared	Date Analyzed	Analyst	Batch ID
TEST METHOD: SW8260B PREP METHOD: SW5030B Test Performed By: AZ0757									
Acetone	<20	20		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Benzene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Bromobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Bromochloromethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Bromodichloromethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Bromoform	<1.0	1.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Bromomethane	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
2-Butanone	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
n-Butylbenzene	<2.5	2.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
sec-Butylbenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
tert-Butylbenzene	<2.5	2.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Carbon disulfide	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Carbon tetrachloride	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Chlorobenzene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Dibromochloromethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Chloroethane	<4.0	4.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Chloroform	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Chloromethane	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
2-Chlorotoluene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
4-Chlorotoluene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2-Dibromo-3-chloropropane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2-Dibromoethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Dibromomethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2-Dichlorobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,3-Dichlorobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,4-Dichlorobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Dichlorodifluoromethane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,1-Dichloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2-Dichloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,1-Dichloroethene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
cis-1,2-Dichloroethene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
trans-1,2-Dichloroethene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2-Dichloropropane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,3-Dichloropropane	<1.0	1.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
2,2-Dichloropropane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,1-Dichloropropene	<1.0	1.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
cis-1,3-Dichloropropene	<1.0	1.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
trans-1,3-Dichloropropene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Ethylbenzene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A

Confidential and Privileged

Trans West Analytical Services
Formerly Columbia Analytical Services

License No. AZ0757/AZM757

CLIENT: Terranext
Work Order: 10060432
Lab ID: 10060432-01
Project Name: WVBA
Project Number: 0310-3154

Client Sample ID: AVB20-0302-1000
Collection Date: 6/28/2010 9:49:00 AM
Matrix: Water

Analyte	Result	PQL	Qual	Units	DF	Date Prepared	Date Analyzed	Analyst	Batch ID
Hexachlorobutadiene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
2-Hexanone	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Iodomethane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Isopropylbenzene	<2.5	2.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
4-Isopropyltoluene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Methylene chloride	<3.0	3.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
4-Methyl-2-pentanone	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Methyl tert-butyl ether	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Naphthalene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
n-Propylbenzene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Styrene	<1.0	1.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,1,1,2-Tetrachloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,1,2,2-Tetrachloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Tetrachloroethene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Toluene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Total Trihalomethanes	<2.5	2.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2,3-Trichlorobenzene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2,4-Trichlorobenzene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,1,1-Trichloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,1,2-Trichloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Trichloroethene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Trichlorofluoromethane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2,3-Trichloropropane	<1.0	1.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2,4-Trimethylbenzene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,3,5-Trimethylbenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Vinyl acetate	<5.0	5.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Vinyl chloride	<0.50	0.50		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
Xylenes, Total	<3.0	3.0		µg/L	1.0	N/A	7/1/10 15:23	NMM	e100701A
4-Bromofluorobenzene(Surrogate)	101	70-135		%REC	1.0	N/A	7/1/10 15:23	NMM	e100701A
Dibromofluoromethane(Surrogate)	102	69-133		%REC	1.0	N/A	7/1/10 15:23	NMM	e100701A
1,2-Dichloroethane-d4(Surrogate)	102	66-139		%REC	1.0	N/A	7/1/10 15:23	NMM	e100701A
Toluene-d8(Surrogate)	107	70-130		%REC	1.0	N/A	7/1/10 15:23	NMM	e100701A

Confidential and Privileged

XENCO Laboratories

Trans West Analytical Services

Formerly Columbia Analytical Services

License No. AZ0757/AZM757

CLIENT: Terranext
Work Order: 10060432
Lab ID: 10060432-02
Project Name: WVBA
Project Number: 0310-3154

Client Sample ID: AVB20-0300-14145
Collection Date: 6/28/2010 9:49:00 AM
Matrix: Water

Analyte	Result	PQL	Qual	Units	DF	Date Prepared	Date Analyzed	Analyst	Batch ID
TEST METHOD: SW8260B PREP METHOD: SW5030B Test Performed By: AZ0757									
Acetone	<20	20		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Benzene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Bromobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Bromochloromethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Bromodichloromethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Bromoform	<1.0	1.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Bromomethane	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
2-Butanone	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
n-Butylbenzene	<2.5	2.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
sec-Butylbenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
tert-Butylbenzene	<2.5	2.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Carbon disulfide	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Carbon tetrachloride	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Chlorobenzene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Dibromochloromethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Chloroethane	<4.0	4.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Chloroform	**3.1**	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Chloromethane	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
2-Chlorotoluene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
4-Chlorotoluene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2-Dibromo-3-chloropropane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2-Dibromoethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Dibromomethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2-Dichlorobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,3-Dichlorobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,4-Dichlorobenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Dichlorodifluoromethane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,1-Dichloroethane	**2.8**	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2-Dichloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,1-Dichloroethene	**5.9**	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
cis-1,2-Dichloroethene	**11**	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
trans-1,2-Dichloroethene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2-Dichloropropane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,3-Dichloropropane	<1.0	1.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
2,2-Dichloropropane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,1-Dichloropropene	<1.0	1.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
cis-1,3-Dichloropropene	<1.0	1.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
trans-1,3-Dichloropropene	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Ethylbenzene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A

Confidential and Privileged

Trans West Analytical Services
Formerly Columbia Analytical Services

License No. AZ0757/AZM757

CLIENT: Terranext
Work Order: 10060432
Lab ID: 10060432-02
Project Name: WVBA
Project Number: 0310-3154

Client Sample ID: AVB20-0300-14145
Collection Date: 6/28/2010 9:49:00 AM
Matrix: Water

Analyte	Result	PQL	Qual	Units	DF	Date Prepared	Date Analyzed	Analyst	Batch ID
Hexachlorobutadiene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
2-Hexanone	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Iodomethane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Isopropylbenzene	<2.5	2.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
4-Isopropyltoluene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Methylene chloride	<3.0	3.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
4-Methyl-2-pentanone	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Methyl tert-butyl ether	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Naphthalene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
n-Propylbenzene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Styrene	<1.0	1.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,1,1,2-Tetrachloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,1,2,2-Tetrachloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Tetrachloroethene	**11**	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Toluene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Total Trihalomethanes	**3.1**	2.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2,3-Trichlorobenzene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2,4-Trichlorobenzene	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,1,1-Trichloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,1,2-Trichloroethane	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Trichloroethene	**77**	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Trichlorofluoromethane	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2,3-Trichloropropane	<1.0	1.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2,4-Trimethylbenzene	<2.0	2.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,3,5-Trimethylbenzene	<1.5	1.5		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Vinyl acetate	<5.0	5.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Vinyl chloride	<0.50	0.50		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
Xylenes, Total	<3.0	3.0		µg/L	1.0	N/A	7/1/10 10:29	NMM	e100701A
4-Bromofluorobenzene(Surrogate)	102	70-135		%REC	1.0	N/A	7/1/10 10:29	NMM	e100701A
Dibromofluoromethane(Surrogate)	102	69-133		%REC	1.0	N/A	7/1/10 10:29	NMM	e100701A
1,2-Dichloroethane-d4(Surrogate)	101	66-139		%REC	1.0	N/A	7/1/10 10:29	NMM	e100701A
Toluene-d8(Surrogate)	105	70-130		%REC	1.0	N/A	7/1/10 10:29	NMM	e100701A

Confidential and Privileged

QC SUMMARY REPORT

XENCO Laboratories
Trans West Analytical Services
Formerly Columbia Analytical Services

Client - Terranext
Work Order - 10060432
Project - WVBA

MB, LCS, LCSD REPORT

SW8260B

Batch ID - e100701A Prep Date - N/A ***Units - µg/L***

Analyte	MB Result	LCS Result	LCSD Result	Spike Value	LCS REC	LCSD REC	Low - High % Limit	% RPD	RPD Limit	MB Qual	LCS Qual	LCSD Qual	Date Analyzed
Acetone	<20	56.3	55.4	40.0	141%	139%	55 - 165	2%	28				07/01/10
Benzene	<0.50	21.6	20.5	20.0	108%	103%	70 - 130	5%	20				07/01/10
Bromobenzene	<1.5	22.5	21.5	20.0	113%	108%	70 - 130	5%	20				07/01/10
Bromochloromethane	<0.50	21.7	21.0	20.0	109%	105%	67 - 125	3%	24				07/01/10
Bromodichloromethane	<0.50	22.3	21.4	20.0	112%	107%	70 - 130	4%	20				07/01/10
Bromoform	<1.0	21.2	21.4	20.0	106%	107%	69 - 130	1%	20				07/01/10
Bromomethane	<5.0	18.7	17.1	20.0	94%	86%	58 - 138	9%	25				07/01/10
2-Butanone	<5.0	47.1	49.3	40.0	118%	123%	58 - 146	5%	27				07/01/10
n-Butylbenzene	<2.5	23.0	22.2	20.0	115%	111%	58 - 128	4%	20				07/01/10
sec-Butylbenzene	<1.5	21.1	20.3	20.0	106%	102%	61 - 133	4%	20				07/01/10
tert-Butylbenzene	<2.5	20.5	19.6	20.0	103%	98%	65 - 128	4%	20				07/01/10
Carbon disulfide	<0.50	19.9	19.3	20.0	100%	97%	59 - 138	3%	22				07/01/10
Carbon tetrachloride	<0.50	20.6	20.1	20.0	103%	101%	57 - 140	2%	21				07/01/10
Chlorobenzene	<0.50	21.2	20.5	20.0	106%	103%	70 - 130	3%	20				07/01/10
Dibromochloromethane	<0.50	22.6	22.3	20.0	113%	112%	70 - 130	1%	20				07/01/10
Chloroethane	<4.0	17.0	16.3	20.0	85%	81%	60 - 146	4%	24				07/01/10
Chloroform	<0.50	20.8	20.0	20.0	104%	100%	66 - 128	4%	24				07/01/10
Chloromethane	<5.0	16.2	14.5	20.0	81%	73%	47 - 144	11%	26				07/01/10
2-Chlorotoluene	<1.5	21.8	20.9	20.0	109%	105%	70 - 130	4%	20				07/01/10
4-Chlorotoluene	<2.0	22.4	21.3	20.0	112%	107%	70 - 130	5%	28				07/01/10
1,2-Dibromo-3-chloropropane	<2.0	21.0	22.2	20.0	105%	111%	60 - 128	6%	21				07/01/10
1,2-Dibromoethane	<0.50	22.4	22.4	20.0	112%	112%	70 - 130	<1%	20				07/01/10
Dibromomethane	<0.50	21.2	21.3	20.0	106%	107%	70 - 130	<1%	23				07/01/10
1,2-Dichlorobenzene	<1.5	22.4	21.6	20.0	112%	108%	70 - 130	4%	20				07/01/10
1,3-Dichlorobenzene	<1.5	21.6	20.7	20.0	108%	104%	70 - 130	4%	20				07/01/10
1,4-Dichlorobenzene	<1.5	20.4	20.1	20.0	102%	101%	70 - 130	1%	20				07/01/10
Dichlorodifluoromethane	<2.0	11.4	10.7	20.0	57%	54%	9 - 134	6%	27				07/01/10
1,1-Dichloroethane	<0.50	20.7	19.7	20.0	104%	99%	66 - 132	5%	20				07/01/10
1,2-Dichloroethane	<0.50	21.3	20.7	20.0	107%	104%	70 - 130	3%	20				07/01/10
1,1-Dichloroethene	<0.50	21.1	20.5	20.0	106%	103%	58 - 144	3%	21				07/01/10
cis-1,2-Dichloroethene	<0.50	21.0	20.7	20.0	105%	104%	67 - 129	1%	24				07/01/10
trans-1,2-Dichloroethene	<0.50	21.7	20.5	20.0	109%	103%	63 - 137	6%	21				07/01/10
1,2-Dichloropropane	<0.50	22.3	21.2	20.0	112%	106%	70 - 130	5%	20				07/01/10
1,3-Dichloropropane	<1.0	21.8	21.7	20.0	109%	109%	70 - 130	<1%	20				07/01/10
2,2-Dichloropropane	<0.50	22.8	22.2	20.0	114%	111%	60 - 141	3%	24				07/01/10
1,1-Dichloropropene	<1.0	22.2	21.8	20.0	111%	109%	64 - 135	2%	20				07/01/10
cis-1,3-Dichloropropene	<1.0	21.6	20.7	20.0	108%	104%	70 - 130	4%	20				07/01/10
trans-1,3-Dichloropropene	<0.50	24.3	23.1	20.0	122%	116%	70 - 130	5%	20				07/01/10
Ethylbenzene	<2.0	22.1	21.5	20.0	111%	108%	70 - 130	3%	20				07/01/10
Hexachlorobutadiene	<5.0	22.3	22.1	20.0	112%	111%	54 - 145	1%	22				07/01/10
2-Hexanone	<5.0	42.7	44.0	40.0	107%	110%	65 - 129	3%	20				07/01/10
Iodomethane	<2.0	15.9	16.5	20.0	80%	83%	68 - 128	4%	22				07/01/10
Isopropylbenzene	<2.5	22.8	22.2	20.0	114%	111%	70 - 130	3%	20				07/01/10
4-Isopropyltoluene	<1.5	21.8	21.1	20.0	109%	106%	67 - 135	3%	20				07/01/10
Methylene chloride	<3.0	19.9	19.2	20.0	100%	96%	61 - 127	4%	20				07/01/10
4-Methyl-2-pentanone	<5.0	40.5	43.1	40.0	101%	108%	67 - 131	6%	21				07/01/10
Methyl tert-butyl ether	<2.0	21.5	21.6	20.0	108%	108%	70 - 130	<1%	21				07/01/10
Naphthalene	<5.0	20.3	20.8	20.0	102%	104%	64 - 133	2%	27				07/01/10

Trans West Analytical Services
Formerly Columbia Analytical Services

QC SUMMARY REPORT

Client - Terranext
Work Order - 10060432
Project - WVBA

MB, LCS, LCSD REPORT

Analyte	MB Result	LCS Result	LCSD Result	Spike Value	LCS REC	LCSD REC	Low - High % Limit	% RPD	RPD Limit	MB Qual	LCS Qual	LCSD Qual	Date Analyzed
n-Propylbenzene	<2.0	22.7	21.8	20.0	114%	109%	65 - 128	4%	20				07/01/10
Styrene	<1.0	21.4	20.5	20.0	107%	103%	70 - 130	4%	20				07/01/10
1,1,1,2-Tetrachloroethane	<0.50	22.4	21.4	20.0	112%	107%	70 - 130	5%	20				07/01/10
1,1,2,2-Tetrachloroethane	<0.50	22.0	22.7	20.0	110%	114%	70 - 130	3%	20				07/01/10
Tetrachloroethene	<0.50	21.6	20.7	20.0	108%	104%	63 - 127	4%	20				07/01/10
Toluene	<2.0	21.2	20.5	20.0	106%	103%	70 - 130	3%	20				07/01/10
Total Trihalomethanes	<2.5	86.9	85.1	80.0	109%	106%	70 - 130	2%	20				07/01/10
1,2,3-Trichlorobenzene	<5.0	21.8	21.4	20.0	109%	107%	66 - 131	2%	27				07/01/10
1,2,4-Trichlorobenzene	<5.0	21.8	21.4	20.0	109%	107%	69 - 127	2%	20				07/01/10
1,1,1-Trichloroethane	<0.50	21.0	20.1	20.0	105%	101%	62 - 133	4%	20				07/01/10
1,1,2-Trichloroethane	<0.50	21.2	21.5	20.0	106%	108%	70 - 130	1%	20				07/01/10
Trichloroethene	<0.50	22.0	21.1	20.0	110%	106%	70 - 130	4%	20				07/01/10
Trichlorofluoromethane	<2.0	19.2	18.0	20.0	96%	90%	45 - 151	6%	22				07/01/10
1,2,3-Trichloropropane	<1.0	21.1	21.9	20.0	106%	110%	70 - 130	4%	20				07/01/10
1,2,4-Trimethylbenzene	<2.0	22.8	21.9	20.0	114%	110%	70 - 130	4%	20				07/01/10
1,3,5-Trimethylbenzene	<1.5	22.1	21.3	20.0	111%	107%	70 - 130	4%	20				07/01/10
Vinyl acetate	<5.0	24.9	25.1	20.0	125%	126%	52 - 142	1%	22				07/01/10
Vinyl chloride	<0.50	18.2	17.4	20.0	91%	87%	43 - 120	4%	25				07/01/10
Xylenes, Total	<3.0	66.1	63.9	60.0	110%	107%	70 - 130	4%	20				07/01/10
4-Bromofluorobenzene	101%	51.1	51.6	50.0	102%	103%	70 - 135						07/01/10
Dibromofluoromethane	99%	49.9	49.7	50.0	100%	99%	69 - 133						07/01/10
1,2-Dichloroethane-d4	98%	48.0	48.4	50.0	96%	97%	66 - 139						07/01/10
Toluene-d8	106%	53.3	53.8	50.0	107%	108%	70 - 130						07/01/10

QC SUMMARY REPORT

XENCO Laboratories
Trans West Analytical Services
Formerly Columbia Analytical Services

Client - Terranext
Work Order - 10060432
Project - WVBA

MS, MSD REPORT

SW8260B

Batch ID - e100701A Prep Date - N/A ***Units - µg/L***
Parent Sample ID - 10060455-08A

Analyte	MS Result	MSD Result	Samp Res	Spike Value	MS REC	MSD REC	Low - High % Limit	% RPD	RPD Limit	MS Qual	MSD Qual	Date Analyzed
Acetone	38.0	40.1	<20	40.0	95%	100%	35 - 160	5%	20			07/01/10
Benzene	20.1	18.9	<0.50	20.0	101%	95%	69 - 130	6%	20			07/01/10
Bromobenzene	20.9	19.9	<1.5	20.0	105%	100%	70 - 130	5%	20			07/01/10
Bromochloromethane	19.5	19.1	<0.50	20.0	98%	96%	63 - 119	2%	22			07/01/10
Bromodichloromethane	24.7	23.5	3.86	20.0	104%	98%	70 - 130	5%	20			07/01/10
Bromoform	19.2	18.6	<1.0	20.0	96%	93%	57 - 121	3%	20			07/01/10
Bromomethane	16.6	14.4	<5.0	20.0	83%	72%	53 - 141	14%	22			07/01/10
2-Butanone	38.8	39.7	<5.0	40.0	97%	99%	46 - 136	2%	22			07/01/10
n-Butylbenzene	22.2	21.1	<2.5	20.0	111%	106%	65 - 127	5%	20			07/01/10
sec-Butylbenzene	20.1	19.3	<1.5	20.0	101%	97%	70 - 130	4%	20			07/01/10
tert-Butylbenzene	19.8	18.6	<2.5	20.0	99%	93%	70 - 130	6%	20			07/01/10
Carbon disulfide	19.3	17.7	<0.50	20.0	97%	89%	58 - 145	9%	28			07/01/10
Carbon tetrachloride	20.9	19.5	<0.50	20.0	105%	98%	60 - 152	7%	20			07/01/10
Chlorobenzene	20.0	18.8	<0.50	20.0	100%	94%	70 - 130	6%	20			07/01/10
Dibromochloromethane	22.5	21.5	1.39	20.0	106%	101%	70 - 130	5%	20			07/01/10
Chloroethane	16.4	15.2	<4.0	20.0	82%	76%	59 - 153	8%	20			07/01/10
Chloroform	25.5	24.8	6.36	20.0	96%	92%	65 - 123	3%	22			07/01/10
Chloromethane	15.1	14.5	<5.0	20.0	76%	73%	47 - 148	4%	22			07/01/10
2-Chlorotoluene	20.5	19.4	<1.5	20.0	103%	97%	70 - 130	6%	20			07/01/10
4-Chlorotoluene	21.0	19.7	<2.0	20.0	105%	99%	70 - 130	6%	20			07/01/10
1,2-Dibromo-3-chloropropane	18.8	18.2	<2.0	20.0	94%	91%	50 - 117	3%	22			07/01/10
1,2-Dibromoethane	20.1	19.7	<0.50	20.0	101%	99%	67 - 117	2%	20			07/01/10
Dibromomethane	19.8	18.9	<0.50	20.0	99%	95%	66 - 115	5%	20			07/01/10
1,2-Dichlorobenzene	20.6	19.7	<1.5	20.0	103%	99%	70 - 130	4%	20			07/01/10
1,3-Dichlorobenzene	20.1	18.9	<1.5	20.0	101%	95%	70 - 130	6%	20			07/01/10
1,4-Dichlorobenzene	19.3	18.7	<1.5	20.0	97%	94%	70 - 130	3%	20			07/01/10
Dichlorodifluoromethane	11.1	9.95	<2.0	20.0	56%	50%	16 - 151	11%	33			07/01/10
1,1-Dichloroethane	19.0	18.3	<0.50	20.0	95%	92%	66 - 129	4%	20			07/01/10
1,2-Dichloroethane	19.5	19.1	<0.50	20.0	98%	96%	64 - 126	2%	20			07/01/10
1,1-Dichloroethene	19.5	19.4	<0.50	20.0	98%	97%	65 - 152	1%	20			07/01/10
cis-1,2-Dichloroethene	20.1	19.2	<0.50	20.0	101%	96%	66 - 126	5%	20			07/01/10
trans-1,2-Dichloroethene	20.1	19.2	<0.50	20.0	101%	96%	66 - 135	5%	20			07/01/10
1,2-Dichloropropane	20.4	19.6	<0.50	20.0	102%	98%	70 - 130	4%	20			07/01/10
1,3-Dichloropropane	20.1	19.3	<1.0	20.0	101%	97%	67 - 115	4%	20			07/01/10
2,2-Dichloropropane	22.0	20.7	<0.50	20.0	110%	104%	62 - 145	6%	20			07/01/10
1,1-Dichloropropene	21.9	20.5	<1.0	20.0	110%	103%	72 - 140	7%	20			07/01/10
cis-1,3-Dichloropropene	19.7	18.4	<1.0	20.0	99%	92%	67 - 122	7%	20			07/01/10
trans-1,3-Dichloropropene	22.0	20.9	<0.50	20.0	110%	105%	70 - 130	5%	20			07/01/10
Ethylbenzene	21.1	19.8	<2.0	20.0	106%	99%	70 - 130	6%	20			07/01/10
Hexachlorobutadiene	22.1	21.3	<5.0	20.0	111%	107%	68 - 143	4%	20			07/01/10
2-Hexanone	34.3	34.0	<5.0	40.0	86%	85%	52 - 122	1%	33			07/01/10
Iodomethane	16.7	15.6	<2.0	20.0	84%	78%	66 - 127	7%	20			07/01/10
Isopropylbenzene	21.7	20.7	<2.5	20.0	109%	104%	70 - 130	5%	20			07/01/10
4-Isopropyltoluene	20.9	19.9	<1.5	20.0	105%	100%	70 - 130	5%	20			07/01/10

QC SUMMARY REPORT

XENCO Laboratories
Trans West Analytical Services
Formerly Columbia Analytical Services

Client - Terranext
Work Order - 10060432
Project - WVBA

MS, MSD REPORT

Analyte	MS Result	MSD Result	Samp Res	Spike Value	MS REC	MSD REC	Low - High % Limit	% RPD	RPD Limit	MS Qual	MSD Qual	Date Analyzed
Methylene chloride	17.8	17.7	<3.0	20.0	89%	89%	59 - 121	1%	20			07/01/10
4-Methyl-2-pentanone	34.8	34.7	<5.0	40.0	87%	87%	53 - 125	<1%	20			07/01/10
Methyl tert-butyl ether	19.2	19.0	<2.0	20.0	96%	95%	65 - 127	1%	20			07/01/10
Naphthalene	18.1	17.8	<5.0	20.0	91%	89%	54 - 129	2%	26			07/01/10
n-Propylbenzene	21.8	20.6	<2.0	20.0	109%	103%	69 - 126	6%	20			07/01/10
Styrene	19.8	18.9	<1.0	20.0	99%	95%	49 - 142	5%	37			07/01/10
1,1,1,2-Tetrachloroethane	20.7	19.7	<0.50	20.0	104%	99%	70 - 130	5%	20			07/01/10
1,1,2,2-Tetrachloroethane	19.5	19.1	<0.50	20.0	98%	96%	64 - 122	2%	20			07/01/10
Tetrachloroethene	21.2	19.9	<0.50	20.0	106%	100%	69 - 130	6%	20			07/01/10
Toluene	20.2	18.9	<2.0	20.0	101%	95%	70 - 130	7%	20			07/01/10
Total Trihalomethanes	91.9	88.4	11.61	80.0	100%	96%	70 - 130	4%	20			07/01/10
1,2,3-Trichlorobenzene	19.7	19.5	<5.0	20.0	99%	98%	61 - 126	1%	24			07/01/10
1,2,4-Trichlorobenzene	20.2	19.8	<5.0	20.0	101%	99%	64 - 123	2%	20			07/01/10
1,1,1-Trichloroethane	20.4	18.8	<0.50	20.0	102%	94%	68 - 136	8%	20			07/01/10
1,1,2-Trichloroethane	19.5	18.7	<0.50	20.0	98%	94%	65 - 112	4%	20			07/01/10
Trichloroethene	21.0	19.4	<0.50	20.0	105%	97%	70 - 130	8%	20			07/01/10
Trichlorofluoromethane	18.4	17.4	<2.0	20.0	92%	87%	53 - 171	6%	20			07/01/10
1,2,3-Trichloropropane	18.5	18.2	<1.0	20.0	93%	91%	58 - 116	2%	20			07/01/10
1,2,4-Trimethylbenzene	21.2	20.0	<2.0	20.0	106%	100%	67 - 128	6%	22			07/01/10
1,3,5-Trimethylbenzene	20.8	19.8	<1.5	20.0	104%	99%	70 - 130	5%	20			07/01/10
Vinyl acetate	21.9	21.6	<5.0	20.0	110%	108%	43 - 133	1%	23			07/01/10
Vinyl chloride	17.7	16.2	<0.50	20.0	89%	81%	46 - 132	9%	21			07/01/10
Xylenes, Total	62.7	58.9	<3.0	60.0	105%	98%	70 - 130	6%	20			07/01/10
4-Bromofluorobenzene	51.4	51.4		50.0	103%	103%	70 - 135					07/01/10
Dibromofluoromethane	49.0	50.0		50.0	98%	100%	69 - 133					07/01/10
1,2-Dichloroethane-d4	47.2	47.7		50.0	94%	95%	66 - 139					07/01/10
Toluene-d8	53.6	53.2		50.0	107%	106%	70 - 130					07/01/10

Container type: VOA

XENCO Laboratories

Sample Receipt Checklist

Client Name: Terranext | Date and Time Received: 6-28-10 1210

Work Order Number: 10060432 | Checked by: [illegible]

Checklist completed by: Tracy [illegible] Date: 6-28-10 | Logged In by: [illegible] Date: 6/28/10

Matrix: Water | Courier Name: Client Xenco _____ | Reviewed by: [illegible] Date: 6/30/10

Shipping container/cooler in good condition?	Yes ☒	No ☐	Not Present ☐	
Custody seals intact on shipping container/cooler?	Yes ☐	No ☐	Not Present ☒	
Custody seals intact on sample bottles?	Yes ☐	No ☐	Not Present ☒	
Chain of custody signed when relinquished and received?	Yes ☒	No ☐	Not Present ☐	
Chain of custody agrees with sample labels?	Yes ☒	No ☐		
Samples in proper container/bottle?	Yes ☒	No ☐		
Sample containers intact?	Yes ☒	No ☐		
Sufficient sample volume for indicated test?	Yes ☒	No ☐		
All samples received within holding time?	Yes ☒	No ☐		
Samples received same day of collection?	Yes ☒	No ☐	Temp: 4.1	Wet Ice Present ☒
Where was the temperature reading taken at?	Sample ☐	Temp Blank ☒	Other:	
VOA Water – VOA vials have zero headspace?	Yes ☒	No ☐	N/A ☐	
Water – Microbiological bottles have ≤ 2.5 cm headspace?	Yes ☐	No ☐	N/A ☒	
Water – All sample pH's acceptable upon receipt?	Yes ☐	No ☐	N/A ☒	Checked by: _____

If **No**, list all samples and bottle types that are not acceptable in Additional Comments section. Also state any correction actions.

Sulfide Water – Bottles have zero headspace?	Yes ☐	No ☐	N/A ☒	(zero headspace ≤ than neck of bottle)
Dissolved Water Analytes – Field Filtered?	Yes ☐	No ☐	N/A ☒	

Are samples received deemed acceptable? Yes ☒ No ☐ If **No** then complete section below

PC Notified Date: _____ Init: _____ PC Init: _____

Client Notified Date: _____ Init: _____ L/M ☐ Date: _____ L/M ☐ Date: _____

Contact Name: _____ Action to take: Analyze ☐ Cancel ☐ Hold ☐ Other:

Changes/Comments made on original COC? Yes ☐ N/A ☐ Init: _____ Date: _____

Changes made in LIMS? Yes ☐ N/A ☐ Init: _____ Date: _____

Additional Comments: _____

Columbia Analytical Services

Water Chain of Custody

Work Order No: 10060432

www.caslab.com •• 3725 E. Atlanta Ave. • Phoenix, AZ 85040 • Ph: (602) 437-0330 •• 3860 S. Palo Verde Rd., Suite 301 • Tucson, AZ 85714 • Ph: (520) 573-1061

Page 1 of 1

Project Manager:	Chuck Gordon	Bill to: (if different)	SAME	Phone:	
Company Name:	Terranext	Company Name:		Special Deliverable Requirements	
Address:	4050 E. Cotton Center Blvd	Address:		Extra surcharges may apply	
City, State ZIP:	Suite 73 PHX, AZ 85040	City, State ZIP:		☐ Level III ☐ Report to MDL	
Phone:	480-496-4100	Email:	cgordon@terranext.net	☐ Level IV ☐ _____	

Special Requirements: ☐ Non Compliant ☒ pdf ☒ EDD ☐ DWAR ☐ NELAP ☐ Project Specific QAPP

Project Name:	WVBA
Project Number:	0310-3154
P.O. Number:	
Sampler's Name:	Roy Ashlock

SAMPLE RECEIPT

Temperature (°C):	4.1	Temp Blank Present:	Y	
Received Intact:	Yes	No	N/A	Wet Ice / Blue Ice
Cooler Custody Seals:	Yes	No	N/A	Total Containers: 4
Sample Custody Seals:	Yes	No	N/A	

TAT: ☒ Routine ☐ Rush - Prelim ☐ Rush - Final

Prelim Due Date:

Final Due Date:

ANALYSIS REQUEST (PLEASE CHECK METHOD NUMBER)

Sample Identification	Matrix	Date Sampled	Time Sampled	Lab ID	Number of Containers	Volatile Organics 8260B ☒ / 624 / 524.2 / 2-CEVE	Acrolein / Acrylonitrile 8260B / 624	Semi-Volatile Organics 8270C / 625	Organochlorine Pesticides 8081A / 608	Oil & Grease (1664-HEM) / TPH (1664-HEM-SGT)	Metals (See Below) Total / Hardness / Ca Hardness	Metals (See Below) Dissolved / TCLP	Total / Dissolved / Hexavalent Cr / Trivalent Cr	Total Cyanide / Amenable (Free) Cyanide	BOD / CBOD	pH / Cond. / Alk / TDS / TSS / TSS-low	Cl / SO4 / F / Ortho-P	NO2 / NO2+NO3 / NO3 (300.0) / NO3 (353.2)	TKN / NH3 / COD / Total-P	Coliform / E. Coli (CFU/MPN) / HPC / Fecal (CFU/MPN)	Comments
AVB20-0302-100	W	6/28/10		1	1	X															
AVB20-0300-145 -14145	W	6/28/10	0949	2	3	X															

Total	200.7 / 6010B:	8RCRA	13PPM	Al	Sb	As	Ba	Be	B	Cd	Ca	Cr	Co	Cu	Fe	Pb	Mg	Mn	Mo	Ni	K	Se	Ag	SiO2	Na	Tl	Sn	V	Zn		
Circle Method(s) and Metal(s) to be analyzed	200.8 / 6020:	Sb	As	Ba	Be	Cd	Cr	Co	Cu	Pb	Mn	Mo	Ni	Se	Ag	Tl	U	Other:												245.1 / 7470A:	Hg

Dissolved / TCLP	200.7 / 6010B:	8RCRA	13PPM	Al	Sb	As	Ba	Be	B	Cd	Ca	Cr	Co	Cu	Fe	Pb	Mg	Mn	Mo	Ni	K	Se	Ag	SiO2	Na	Tl	Sn	V	Zn		
Circle Method(s) and Metal(s) to be analyzed	200.8 / 6020:	Sb	As	Ba	Be	Cd	Cr	Co	Cu	Pb	Mn	Mo	Ni	Se	Ag	Tl	U	Other:												245.1 / 7470A:	Hg

	Relinquished by: (Signature)	(Print Name)	Received by: (Signature)	(Print Name)	Date/Time
1	[signature]	Roy Ashlock	[signature]	Byron Schneid	6/28/10 @ 11:45A
2	[signature]	Byron Schneid	[signature]	Tracey Mitchell	6-28-10 12/0
3					

System Name: PWS #:

Date and Time samples arrived at reporting lab: Phoenix / Tucson

Revised : 8/6/2009